

Published in final edited form as:

Nat Methods. 2011 March ; 8(3): 246–249. doi:10.1038/nmeth.1558.

Micropilot: automation of fluorescence microscopy–based imaging for systems biology

Christian Conrad¹, Annelie Wünsche², Tze Heng Tan², Jutta Bulkescher¹, Frank Sieckmann³, Fatima Verissimo², Arthur Edelstein⁴, Thomas Walter², Urban Liebel^{2,5}, Rainer Pepperkok^{1,2}, and Jan Ellenberg²

¹Advanced Light Microscopy Facility, European Molecular Biology Laboratory, Heidelberg, Germany.

²Cell Biology and Biophysics Unit, European Molecular Biology Laboratory, Heidelberg, Germany.

³Leica Microsystems GmbH, Mannheim, Germany.

⁴Department of Cellular and Molecular Pharmacology, University of California, San Francisco, California, USA.

Abstract

Quantitative microscopy relies on imaging of large cell numbers but is often hampered by time-consuming manual selection of specific cells. The ‘Micropilot’ software automatically detects cells of interest and launches complex imaging experiments including three-dimensional multicolor time-lapse or fluorescence recovery after photobleaching in live cells. In three independent experimental setups this allowed us to statistically analyze biological processes in detail and is thus a powerful tool for systems biology.

The fluorescence microscope has developed into a versatile quantitative measuring device in cell biology. Advanced imaging techniques including four-dimensional imaging and fluorescence recovery after photobleaching (FRAP) can be used to visualize and analyze complex biological processes in cells, organelles and subcellular compartments^{1,2}. However, quantification with statistical significance requires sampling high numbers of events. This poses a challenging problem for the quantification of rare cellular events.

In principle, computational image analysis can be used to detect specific cells in microscopy images automatically. Several open-source packages for postacquisition image processing of

© 2011 Nature America, Inc. All rights reserved.

Correspondence should be addressed to R.P. (pepperko@embl.de) or J.E. (jan.ellenberg@embl.de).

⁵Present address: Institute of Toxicology and Genetics, Karlsruhe Institute of Technology, Eggenstein-Leopoldshafen, Germany.

Note: Supplementary information is available on the Nature Methods website.

AUTHOR CONTRIBUTIONS

C.C. developed the ‘Micropilot’ software and drafted the manuscript. A.W. developed the Visual Basic for Applications macro and performed and analyzed the automatic FRAP experiments. T.H.T. developed the feature selection, extended classification to multiple channels and acquired and analyzed the ERES images. F.V. performed the ERES experiments. J.B. performed and analyzed the spindle length experiments. F.S. and U.L. developed the computer-aided microscopy interface and set up software prototypes. A.E. developed the communication of µManager with Micropilot. T.W. helped with image processing and object feature design. R.P. supervised the project. J.E. supervised the project and revised the manuscript.

COMPETING FINANCIAL INTERESTS

The authors declare competing financial interests: details accompany the full-text HTML version of the paper at <http://www.nature.com/naturemethods/>.

Reprints and permissions information is available online at <http://npg.nature.com/reprintsandpermissions/>.

cell-based assays already include classification by machine learning for cell recognition. Examples are CellProfiler Analyst/CellClassifier^{3,4} and EBImage⁵, which provide workflows for fixed cell images as well as CellCognition⁶, a package dedicated to the classification of large time-lapse datasets with a fast image-processing engine. Several commercial tools, for example, Acapella from Perkin-Elmer or ScanR from Olympus, enable analysis of a few generally applicable cell-based assays but lack the extensibility to new algorithms that open-source packages have. In contrast, not only dedicated screening platforms but also state-of-the-art research microscopes have become fully motorized and computer controlled either by microscope vendor or open-source software packages such as μ Manager. But there has been no software solution to leverage the power of machine learning based cell recognition to automatically control the whole repertoire of fluorescence-based imaging assays.

To fill this gap, we developed the Micropilot software that effectively replaces a human expert at the microscope to identify cells of interest and automate even complex fluorescence microscopy-based imaging assays (Supplementary Software). Micropilot provides a machine learning-based module that can be rapidly trained by the user to automatically identify desired cell states during an unattended, fast, low-resolution prescanning mode. Upon identification of a cell of interest Micropilot switches to a complex imaging mode that automatically runs desired complex high-resolution and high-information-content imaging assays such as high-resolution three-dimensional (3D) time-lapse imaging or FRAP. After completion of the desired complex imaging procedure, Micropilot returns to the low-resolution prescanning mode to identify additional cells of interest (Fig. 1a). We used Lab View for the main Micropilot program and C for the image analysis routines. We tested Micropilot on the Microsoft Windows platform.

Micropilot relies on a robust communication channel to read the low-resolution prescan microscope images, interrupt the scan to analyze them and reconfigure the microscope for the desired complex imaging procedure (Fig. 1b,c). Pseudocode examples are listed in Supplementary Table 1. The hardware prerequisites for such automatic imaging systems are (i) motorization of the microscope stage, (ii) automatic changing of objectives and/or laser scanner zoom, and (iii) switching of fluorescence filter and/or laser lines. The microscope system should also provide multipositioning or grid scanning along with autofocus capabilities. Micropilot can remotely control any microscope offering interfaces to this specification. We implemented Micropilot for Leica SP5, μ Manager, Olympus ScanR, PerkinElmer Ultra View ERS and Zeiss 510. Certainly, μ Manager as an open-source microscope control software allocates most freedom for users in terms of further developments.

To test the power of Micropilot, we applied this platform to three different fluorescence microscopy-based assays, focusing on the biological process of cell division for which low-resolution imaging data from functional genomic studies have recently become available^{7,8}. We aimed to identify automatically the two most transient mitotic stages, prophase and anaphase, in low-resolution prescan images and wanted to assay three different biological processes: (i) the biogenesis of endoplasmic reticulum exit sites (ERESs) after mitosis using high-resolution immunofluorescence, (ii) the assembly of the microtubule spindle using 3D time-lapse imaging and (iii) the binding of chromobox homolog 1 to condensing chromosomes using FRAP. To test the hardware flexibility of the software, we used two different microscopy systems for these assays.

ERESs disappear during metaphase in mitosis and reassemble at the telophase or cytokinesis transition⁹. To understand ERES biogenesis quantitatively, we trained the Micropilot classifier to recognize mitotic cells by their chromosome configuration in the DNA

(Hoechst) channel and select those that showed quantifiable expression of YFP-tagged SEC31, a COPII component and ERES marker (Fig. 2a). The trained classifier provided 92% sensitivity (Supplementary Fig. 1a) with a positive predictive value (PPV) of 88% (Fig. 2b). For three microscope slides containing HeLa-Kyoto cells, we acquired 2,700 image fields at low resolution and detected 173 high-resolution confocal 3D stacks (each comprising 50 slices) of anaphase and later mitotic stages (Fig. 2c) during three unattended overnight scans with only 1.5 h combined hands-on time. After manual quality inspection of the image data, we quantified the number of ERESs in 91 of the high-resolution 3D stacks. This revealed that the average number of ERESs exponentially increased with the volume increase of the nuclei (Fig. 2d) but showed no clear correlation with the increase in inter-nuclei distance (Supplementary Fig. 1b). As nuclear volume marks the progression of cells from anaphase to cytokinesis, these data are consistent with earlier work⁹ suggesting ERES biogenesis after mitosis to be most efficient at the telophase or cytokinesis transition. In the future, extending the experiments we performed with additional markers of ERES composition and systematic microscopy-based quantification enabled by Micropilot will shed more light on the mechanism of postmitotic ERES biogenesis.

The assembly of the mitotic microtubule spindle is a highly dynamic process, which starts upon mitotic entry in prophase and is a key prerequisite for genome segregation. We aimed to automatically image prophase cells and the first appearance of microtubule asters in four dimensions, which enables dynamic measurement of structural parameters of the forming spindle such as its length. We trained the Micropilot classifier to identify prophase HeLa cells based on the condensing chromatin (histone H2B tagged with mCherry, abbreviated H2B-mCherry), and subsequently recorded fluorescently labeled microtubules (EGFP-tagged α -tubulin, abbreviated EGFP-tubulin) and chromosomes in two-color, 3D time-lapse sequences of live cells at high spatial and temporal resolution. We fed 3,644 fluorescently labeled H2B-mCherry cells into the Micropilot multiclass support vector machine (SVM) classifier to recognize six major mitotic phases (Supplementary Fig. 2a). The resulting classifier achieved 97% mean sensitivity and 87% mean PPV (Supplementary Fig. 2b) in good accordance with other mitotic classifications^{10,11} and could be used without retraining to control a second confocal microscope from the same manufacturer with very similar prophase PPV of 85% (Supplementary Fig. 2c). These findings indicate good generalization properties of the classifier that is not dramatically affected by hardware variations of the two similar confocal microscopes or biological sample variation over 6 months. In 15 unattended overnight runs, more than 120 high-resolution, 3D time-lapse mitosis movies were generated, instead of an expected 2 months of full-time manual work by an experienced microscopist (Fig. 2e and Supplementary Video 1). Quantitative analysis of the movies revealed that the mean pole-to-pole distance of the metaphase spindle of HeLa cells was $15.6 \pm 1.6 \mu\text{m}$ (\pm s.d.) and increased to $25.3 \pm 1.7 \mu\text{m}$ in telophase (Fig. 2e,f), which is consistent with previous measurements¹². Exploiting this automatic quantitative assay for spindle assembly, we analyzed the cells in which expression of the essential centromere protein CENP-E had been knocked down by siRNA. Automatic sampling of 71 CENPE-depleted cells arrested in prophase or prometaphase¹³, revealed three different subpopulations of spindle lengths, with mean pole-to-pole distances of $15.5 \pm 3.2 \mu\text{m}$, which correlated with the number of unaligned chromosomes at the poles (Fig. 2g,h). The good statistical sampling of prophase or prometaphase stages enabled by Micropilot thus revealed that depletion of CENP-E activity by RNAi led to different metastable spindle defects rather than a continuous spectrum of abnormal spindle lengths.

Condensation of mitotic chromosomes is a dynamic structural reorganization of the genome in mitotic prophase and is poorly understood molecularly. To define candidate proteins for this process their binding to condensing chromosomes needs to be assayed. As an example candidate we focused on chromobox homolog 1 (CBX1 or heterochromatin protein 1 beta

(HP1 β). In interphase, CBX1 binds in heterochromatic areas of chromosomes¹⁴, but it is largely released from chromosomes in metaphase and freely diffuses in the cell¹⁵. To analyze CBX1 binding to condensing chromosomes, we performed automatic FRAP analysis on prophase and interphase nuclei. We trained the Micropilot classifier to identify prophase 3T3 cells based on the condensing chromatin (H2B-mCherry), and subsequently performed automatic FRAP experiments on CBX1-EGFP by photobleaching half of the nucleus and following the recovery of CBX1 fluorescence during 1 min (Fig. 3a and Supplementary Videos 2–4). The binary prophase classifier had 92% mean sensitivity and 95% mean PPV on untrained images (Supplementary Fig. 3a). In four unattended overnight runs we bleached 115 interphase and 117 prophase cells. The manual selection of the rare prophase stages would have required a full month of work by an expert microscopist. The recovery curves of 50 interphase cells exhibited a characteristic halftime of recovery ($t_{1/2}$) of 6.7 ± 1.4 s (\pm s.d.), resulting from the exchange of the heterochromatin-bound pool (Fig. 3b). The $t_{1/2}$ values of 64 prophase cells were distributed in two distinct populations with distinct kinetics (Supplementary Fig. 3b) suggesting that CBX1-chromatin interactions undergo discrete changes during prophase. By analyzing the chromosome condensation state in the two populations, we found that the two clusters of recovery halftimes corresponded to early and late condensation stages (Supplementary Fig. 3b). In early prophase CBX1 recovered significantly slower than in interphase with a mean $t_{1/2}$ of 7.8 ± 0.9 s ($P = 0.0013$), indicating that its binding is increased during condensation. By contrast, in late prophase, when chromosome condensation was nearly complete, CBX1 was largely unbound and recovered significantly faster than in interphase with a mean $t_{1/2}$ of 4.2 ± 1.1 s ($P < 0.0001$) (Fig. 3c). The good statistical sampling of prophase stages enabled by Micropilot revealed that CBX1 binding to condensing chromosomes is transiently elevated in early prophase before its release at the end of prophase, suggesting a possible function in the condensation process.

To create quantitative data suitable for systems biology at the cellular level of biology, intelligent automation will be crucial to increase throughput and provide statistical relevance of the measurements of very different cellular processes. Micropilot automation reduces bias in cell selection and liberates cell biologists from the tedious work of repetitive manual data generation. Micropilot is broadly applicable in cell biology to carry out a host of quantitative imaging assays in a systematic fashion as required to characterize a large number of genes or perturbations such as in RNAi studies or drug screens. It can be adapted to virtually any imaging system that allows automation and online control based on the results of image classification by machine vision. It will be straightforward in the future to use Micropilot to automate additional high-resolution assays beyond multicolor high-resolution confocal 3D stacks, time-lapse movies and FRAP, such as fluorescence lifetime imaging microscopy and fluorescence cross-correlation spectroscopy.

METHODS

Methods and any associated references are available in the online version of the paper at <http://www.nature.com/naturemethods/>.

Supplementary Material

Refer to Web version on PubMed Central for supplementary material.

Acknowledgments

This study was technically supported by the use of the European Molecular Biology Laboratory Advanced Light Microscopy Facility and Information Technology Service Unit. We acknowledge C. Chapuis for the *CBX1-EGFP* BAC cell line expressing H2B-mCherry. This project was funded by grants to J.E. and R.P. (within the MitoCheck,

MitoSys and Systems Microscopy consortia) by the European Commission (LSHG-CT-2004-503464, FP7/2007-2013-241548 and FP7/2007-2013-258068), to R.P. by the Landesstiftung Baden-Württemberg in the framework of the research program RNS/RNAi and within the Nationales Genomforschungsnetz-Plus consortium IG-CSG (01GS0865). T.H.T. is financed by the Deutsche Forschungsgemeinschaft Graduiertenkolleg GRK118.

References

1. Pepperkok R, Ellenberg J. *Nat. Rev. Mol. Cell Biol.* 2006; 7:690–696. [PubMed: 16850035]
2. Conrad C, Gerlich DW. *J. Cell Biol.* 2010; 188:453–461. [PubMed: 20176920]
3. Ramo P, Sacher R, Snijder B, Begemann B, Pelkmans L. *Bioinformatics.* 2009; 25:3028–3030. [PubMed: 19729371]
4. Jones TR, et al. *Proc. Natl. Acad. Sci. USA.* 2009; 106:1826–1831. [PubMed: 19188593]
5. Pau G, Fuchs F, Sklyar O, Boutros M, Huber W. *Bioinformatics.* 2010; 26:979–981. [PubMed: 20338898]
6. Held M, et al. *Nat. Methods.* 2010; 7:747–754. [PubMed: 20693996]
7. Hutchins JR, et al. *Science.* 2010; 328:593–599. [PubMed: 20360068]
8. Neumann B, et al. *Nature.* 2010; 464:721–727. [PubMed: 20360735]
9. Hammond AT, Glick BS. *Mol. Biol. Cell.* 2000; 11:3013–3030. [PubMed: 10982397]
10. Neumann B, et al. *Nat. Methods.* 2006; 3:385–390. [PubMed: 16628209]
11. Wang M, et al. *Bioinformatics.* 2008; 24:94–101. [PubMed: 17989093]
12. Bird AW, Hyman AA. *J. Cell Biol.* 2008; 182:289–300. [PubMed: 18663142]
13. Schaar BT, Chan GK, Maddox P, Salmon ED, Yen TJ. *J. Cell Biol.* 1997; 139:1373–1382. [PubMed: 9396744]
14. Muller KP, et al. *Biophys. J.* 2009; 97:2876–2885. [PubMed: 19948116]
15. Schmiedeberg L, Weisshart K, Diekmann S, Meyer Zu Hoerste G, Hemmerich P. *Mol. Biol. Cell.* 2004; 15:2819–2833. [PubMed: 15064352]




Figure 1. Schematic workflow of Micropilot. **(a)** After autofocussing different positions to find the best focal plane (yellow frame), low-resolution prescan images (optionally maximum z-dimension projections, gray frames) are presented to the automatic classification. If a cell is selected, a complex imaging protocol is executed; otherwise the system continues to prescan. After completion of the complex imaging protocol, the system loops back to prescan mode, continuing at the sample position where it stopped for the complex imaging mode. **(b)** Communication steps executed by the different microscope systems (red outlines) and the Micropilot software (blue outlines). The microscope sends the image path either via windows registry or socket interface to Micropilot. In the synchronous modes, each positive classification launches the complex imaging mode. In the asynchronous mode, microscope and Micropilot send and receive messages via transmission control protocol or internet protocol (TCP/IP), allowing classification of several different positions before launching the complex imaging protocol for a list of positions. **(c)** After reading the low-resolution image, Micropilot segments, extracts the feature set per object and classifies the cells during scanning to return eventually the positions of interest. After the criteria are met (time or number of positions) Micropilot deploys the complex imaging and the microscope switches back to prescan mode **(a)**.




Figure 2.

Assays of *SEC31* and *H2B-tubulin* HeLa cells. **(a)** Examples for Hoechst-labeled (blue; DNA label) and SEC31-labeled (green) cells representing null or artifact and anaphase or telophase cells (insets, close-up images). Scale bars, 10 μm . **(b)** Confusion matrix of the prediction shows true positives (TP) horizontally against the predicted class vertically for cells. At edges the total numbers of the cells are given (overall total, 10,793 cells) corresponding to $\text{PPV} = \text{TP} / (\text{TP} + \text{false positives})$ and $\text{sensitivity} = \text{TP} / (\text{TP} + \text{false negatives})$. **(c)** Examples of null or artifact (left) and anaphase or telophase (right) cells stained with Hoechst (blue) and ERES spot (green) (50 slices of 0.2 μm). Scale bar, 10 μm . **(d)** Number of ERES spots of 91 anaphase cells to late-telophase cells plotted versus volume of nuclei, with exponential fit plotted. Red and blue data points correspond to the nuclei in the left and right images in **c**, respectively. **(e)** Example of negative control experiment (time resolution, 3 min; 30 slices of 1 μm ; maximum projections) started after prophase recognition. Times indicated are after prophase recognition. Scale bar, 10 μm . **(f)** Spindle lengths after treatment with scrambled siRNA. **(g)** Example images after treatment with siRNA to *CENPE*, showing centrosome poles (arrows; left) for the first recognizable metaphase (acquisition as in **e**). Scale bar, 10 μm . **(h)** Normal mixture modeling of pole-pole distances in metaphase from 71 movies after treatment with siRNA to *CENPE* resulted in three distributions, which are shown as colored curves.




Figure 3.

Examples and measurements of automatic FRAP on *CBX1-EGFP* cells. **(a)** After the automatic selection of an interphase or prophase cell with a trained prophase SVM classifier, a prebleached image was taken, followed by bleaching of the predefined upper half of the nucleus and subsequent time-lapse imaging with 2-s time resolution for 60 s (values in the lower images indicate time relative to bleaching). Scale bar, 5 μ m. **(b)** Normalized intensities for CBX1-EGFP measured during fluorescence relaxation after photobleaching in interphase and prophase cells. We measured, normalized, averaged and plotted over time fluorescence intensities in bleached region of the nucleus. Error bars, s. **(c)** Recovery rates as box plots for interphase and prophase cell populations.